



राष्ट्रीय वनस्पति स्वास्थ्य प्रबंधन संस्थान
National Institute of Plant Health Management
कृषि एवं सहकारिता विभाग, कृषि एवं किसान कल्याण मंत्रालय, भारत सरकार
Department of Agriculture & Farmers Welfare,
Ministry of Agriculture & Farmers Welfare, Government of India



Telephone: 9140-24015374
E-mail: niphm@nic.in
Tele-Fax: 9140-24015346

Rajendra Nagar,
Hyderabad – 500 030
<http://niphm.gov.in>

F. No. 16/PMD.PROC./HPLC-PDA/2021-22/15

Date:- 30/12/2021

CORRIGENDUM

Sub: Corrigendum to Tender-Technical Specification - Reg.
Ref: Tender ID No. 2021_DACO_663322_1

This has reference to Tender ID Number: 2021_DACO_663322_1 dt.15.12.2021 in connection with the procurement of the equipment viz., 'High Performance Liquid Chromatography Unit Coupled with Photo Diode Array Detector (HPLC-PDA), with AMC Service (without spares only service support) for 5 Years, after Warranty Period of minimum 3 Years'

2. Bidders may note the amendment to Technical Specification as under.

Clause No of the approved specification	Name of Item	Existing Technical specification	Change in Technical specification
Please refer to clause-3 of page-15 under Section-VI: Technical Specifications	'High Performance Liquid Chromatography Unit Coupled with Photo Diode Array Detector (HPLC-PDA), with AMC Service (without spares only service support) for 5 Years, after Warranty Period of minimum 3 Years'	Column Oven (optional): i. Column oven to accommodate up to two 25- 30 cm Columns , preferably a microprocessor based, conforming to ISO/IEC: 17025 with the NABL accredited lab certificate. ii. Should have a range of ambient to 65°C with 1°C increments iii. Should have indicators/monitors to know the status of the instrument	This subject clause may be treated as omitted from the Tender document. Hence, agencies are requested to quote the price for the High Performance Liquid Chromatography Unit coupled with Photo Diode Array Detector without a column oven.

It has been decided to revise the last date of bid submission and the date of bid opening. The revised dates are given below:

Sl. No.	Existing Dates	Revised Dates
1.	End Date : 04/01/2022 at 03.00 pm. Opening Date : 05/01/2022 at 04.00 p.m.	End Date : 07/01/2022 at 03.00 pm. Opening Date : 08/01/2022 at 04.00 p.m.

The above corrigendum to Tender shall be part of the Bid Document for all purposes. All other terms and conditions in the tender document remain the same.

Registrar I/c